

Better wound care begins with Keystone.

Our Wound Care Program is designed with the clinician and patient in mind. We work closely with health care teams and wound care specialists to address wound care's unique challenges to support faster wound healing.

The Cork Nisus Negative Pressure Wound Therapy Pump

is part of a complete wound care system. Two essential components that actively work together to promote wound healing are the Nisus NPWT **Canisters** and the NPWT **Wound Dressing Kit**. Also included with your NPWT system are the Cork **battery charger** and **carrying bag**.

Once the device is delivered, notify a member of your healthcare team, to coordinate the best time for the Nisus system to be applied.



Things you Need to Know About Your Nisus NPWT Device

The Nisus pump will be worn 24 hours a day throughout your treatment and should be kept in the black carrying bag provided.

Do not allow the Nisus pump to get wet. Disconnect the tubing from the canister if you take a bath or shower. Do not submerge the pump in water.

If the device gets wet, unplug it immediately.

Keep the Nisus plugged into a power outlet whenever possible to keep the battery fully charged. Expected 24 hour battery life between charges.

Keep the Nisus upright to avoid a false "Canister Full" alert.

Always keep the Nisus turned on unless there is bleeding from the wound or instructed by your health care professional.

Do not change the settings on the Nisus unless instructed by your health care professional.

Notify the treating practitioner for an unexpected increase in fluid from the wound.

Canisters are single-use only and must be changed at a minimum weekly or when the canister is full.

It's normal to hear the pump turning on and off as it re-pressurizes your wound to keep it at appropriate settings.

If the NPWT device is off for two or more hours, remove the dressing and foam and place a moist dressing on the wound. Once this is done, notify your medical provider immediately.

Important Contact Information: If you have a MEDICAL PROBLEM, please call your clinic or physician.

If you have a PUMP PROBLEM, please call the Keystone support line at 601-707-9727



How do I return my pump?

Your healthcare professional will determine when your wound therapy has ended. Our clinical coordinator will contact you at that time to provide instructions for returning the pump and battery charger.

Do not return the pump to a clinic or other health care providers. Keystone will send a prepaid return shipping box to your home so the **pump and battery charger** will be returned directly to Keystone. Simply follow the instructions that come with the box.